

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	0.160 °C	1.1 °C	0.940 °C	± 1.0 °C	± 4.0E-01 °C
10 °C	10.348 °C	11.1 °C	0.752 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.693 %HR	27.9 %HR	-2.793 %HR	± 1.0 %HR	± 1.9E+00 %HR
60 %HR	59.793 %HR	55.1 %HR	-4.693 %HR	± 5.0 %HR	± 1.9E+00 %HR