

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	0.160 °C	0.3 °C	0.140 °C	± 1.0 °C	± 4.0E-01 °C
10 °C	10.348 °C	10.6 °C	0.252 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.693 %HR	28.3 %HR	-2.393 %HR	± 5.0 %HR	± 1.9E+00 %HR
60 %HR	59.793 %HR	56.3 %HR	-3.493 %HR	± 5.0 %HR	± 1.9E+00 %HR