

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	0.160 °C	0.3 °C	0.140 °C	± 1.0 °C	± 4.0E-01 °C
10 °C	10.348 °C	10.5 °C	0.152 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.693 %HR	30.1 %HR	-0.593 %HR	± 1.0 %HR	± 1.9E+00 %HR
60 %HR	59.793 %HR	57.7 %HR	-2.093 %HR	± 5.0 %HR	± 1.9E+00 %HR