

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	0.160 °C	0.7 °C	0.540 °C	± 1.0 °C	± 4.0E-01 °C
15 °C	15.061 °C	15.5 °C	0.439 °C	± 1.0 °C	± 4.0E-01 °C
35 °C	35.046 °C	35.4 °C	0.354 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.693 %HR	28.0 %HR	-2.693 %HR	± 5.0 %HR	± 1.9E+00 %HR
60 %HR	59.793 %HR	55.9 %HR	-3.893 %HR	± 5.0 %HR	± 1.9E+00 %HR