

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
25 °C	24.828 °C	24.7 °C	-0.128 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.824 %HR	32.5 %HR	2.676 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.621 %HR	51.2 %HR	0.579 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	74.824 %HR	72.1 %HR	-2.724 %HR	± 3.0 %HR	± 1.9E+00 %HR