

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.143 °C	30.4 °C	0.257 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.243 %HR	33 %HR	2.757 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	51.388 %HR	53 %HR	1.612 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.151 %HR	75 %HR	4.849 %HR	± 5.0 %HR	± 1.9E+00 %HR