

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
25 °C	25.169 °C	25.0 °C	-0.169 °C	± 1.0 °C	± 4.0E-01 °C
50 %HR	50.468 %HR	50.1 %HR	-0.368 %HR	± 3.0 %HR	± 1.9E+00 %HR