

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
52.65 μm	52.43 μm	-0.22 μm	± 2.5 μm	±	5.8E-02 μm
155.23 μm	155.6 μm	0.37 μm	± 2.5 μm	±	6.2E-02 μm
257.12 μm	257.8 μm	0.68 μm	± 2.5 μm	±	7.0E-02 μm
511.71 μm	511.0 μm	-0.71 μm	± 2.5 μm	±	9.7E-02 μm