

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(M.P.E)</i>	Incertidumbre <i>(Uncertainty)</i>	
99.54257 µl	100 µl	0.45743 µl	± 0.8 µl	±	5.1E-02 µl