

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(M.P.E)</i>	Incertidumbre <i>(Uncertainty)</i>	
1,000.23743 µl	1000 µl	-0.23743 µl	± 8 µl	±	7.8E-02 µl