

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(M.P.E)</i>	Incertidumbre <i>(Uncertainty)</i>	
9.95426 µl	10 µl	0.04574 µl	± 0.12 µl	±	5.1E-02 µl
50.02190 µl	50 µl	-0.02190 µl	± 0.5 µl	±	5.1E-02 µl
100.14404 µl	100 µl	-0.14404 µl	± 0.8 µl	±	5.1E-02 µl