

Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	E.M.P (<i>M.P.E</i>)	Incertidumbre (<i>Uncertainty</i>)	
1.97481 µl	#2 µl	0.02519 µl	± 0.1 µl	±	5.1E-02 µl
4.99217 µl	#5 µl	0.00783 µl	± 0.1 µl	±	5.5E-02 µl
9.95426 µl	10 µl	0.04574 µl	± 0.1 µl	±	5.1E-02 µl