

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (M.P.E)	Incertidumbre (Uncertainty)	
1.00234 ml	1 ml	-0.00234 ml	± 0.008 ml	±	7.8E-02 µl
4.99978 ml	5 ml	0.00022 ml	± 0.04 ml	±	3.0E-01 µl
10.00739 ml	10 ml	-0.00739 ml	± 0.06 ml	±	5.9E-01 µl