

Patrón (Standard)	UBP (UIT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
5 °	5 °	0 °	± 1 °	±	5.8E-01 °
10 °	10 °	0 °	± 1 °	±	5.8E-01 °
20 °	20 °	0 °	± 1 °	±	5.8E-01 °
40 °	40 °	0 °	± 1 °	±	5.8E-01 °
60 °	60 °	0 °	± 1 °	±	5.8E-01 °
90 °	90 °	0 °	± 1 °	±	5.8E-01 °
100 °	100 °	0 °	± 1 °	±	5.8E-01 °
110 °	110 °	0 °	± 1 °	±	5.8E-01 °
116 °	116 °	0 °	± 1 °	±	5.8E-01 °