

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.069 °C	-24.7 °C	0.369 °C	± 0.5 °C	± 4.0E-01 °C
0 °C	-0.281 °C	0.2 °C	0.481 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	7.741 °C	8.2 °C	0.459 °C	± 0.5 °C	± 4.0E-01 °C