

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
12 °C	11.973 °C	13.1 °C	1.127 °C	± 0.5 °C	± 4.0E-01 °C
27 °C	26.793 °C	28.1 °C	1.307 °C	± 0.5 °C	± 4.0E-01 °C
42 °C	41.805 °C	42.8 °C	0.995 °C	± 0.5 °C	± 4.0E-01 °C
52 %HR	50.601 %HR	52.8 %HR	2.199 %HR	± 3.0 %HR	± 1.9E+00 %HR
72 %HR	72.186 %HR	71.1 %HR	-1.086 %HR	± 3.0 %HR	± 1.9E+00 %HR
# 92 %HR	92.390 %HR	92.4 %HR	0.010 %HR	± 3.0 %HR	± 1.9E+00 %HR