

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
12 °C	11.973 °C	12.3 °C	0.327 °C	± 0.5 °C	± 4.0E-01 °C
27 °C	26.793 °C	27.1 °C	0.307 °C	± 0.5 °C	± 4.0E-01 °C
42 °C	41.805 °C	41.6 °C	-0.205 °C	± 0.5 °C	± 4.0E-01 °C
52 %HR	50.601 %HR	53.6 %HR	2.999 %HR	± 3.0 %HR	± 1.9E+00 %HR
72 %HR	72.186 %HR	75.1 %HR	2.914 %HR	± 3.0 %HR	± 1.9E+00 %HR
# 92 %HR	92.390 %HR	89.3 %HR	-3.090 %HR	± 3.0 %HR	± 1.9E+00 %HR