

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	0.010 °C	0.1 °C	0.090 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.144 °C	5.3 °C	0.156 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.127 °C	10.6 °C	0.473 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.660 %HR	33.2 %HR	3.540 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.149 %HR	53.2 %HR	4.051 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.366 %HR	74.7 %HR	4.334 %HR	± 3.0 %HR	± 1.9E+00 %HR