

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	0.010 °C	-0.2 °C	-0.210 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.144 °C	4.9 °C	-0.244 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.127 °C	10.2 °C	0.073 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.660 %HR	32.6 %HR	2.940 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.149 %HR	48.2 %HR	-0.949 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.366 %HR	67.6 %HR	-2.766 %HR	± 3.0 %HR	± 1.9E+00 %HR