

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
25 °C	24.026 °C	23.8 °C	-0.226 °C	± 1.0 °C	± 4.0E-01 °C
45 %HR	43.351 %HR	40.7 %HR	-2.651 %HR	± 6.0 %HR	± 1.9E+00 %HR