

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.552 °C	15.9 °C	0.348 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.199 °C	25.5 °C	0.301 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.168 °C	32.1 °C	-0.068 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	30.0 %HR	-0.570 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.141 %HR	69.8 %HR	-0.341 %HR	± 3.0 %HR	± 1.9E+00 %HR