

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	0.130 °C	0.3 °C	0.170 °C	± 0.45 °C	± 4.0E-01 °C
5 °C	5.003 °C	5.2 °C	0.197 °C	± 0.45 °C	± 4.0E-01 °C
10 °C	10.318 °C	10.4 °C	0.082 °C	± 0.45 °C	± 4.0E-01 °C
15 °C	15.131 °C	15.2 °C	0.069 °C	± 0.45 °C	± 4.0E-01 °C
25 °C	24.808 °C	24.7 °C	-0.108 °C	± 0.45 °C	± 4.0E-01 °C
32 °C	31.797 °C	31.4 °C	-0.397 °C	± 0.45 °C	± 4.0E-01 °C
30 %HR	30.376 %HR	33.9 %HR	3.524 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	69.814 %HR	73.1 %HR	3.286 %HR	± 3.0 %HR	± 1.9E+00 %HR