

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
10 °C	10.338 °C	11.2 °C	0.862 °C	± 1.0 °C	± 4.0E-01 °C
23 °C	22.913 °C	23.4 °C	0.487 °C	± 1.0 °C	± 4.0E-01 °C
40 °C	39.839 °C	39.5 °C	-0.339 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.458 %HR	34 %HR	3.542 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.069 %HR	49 %HR	-1.069 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	74.998 %HR	73 %HR	-1.998 %HR	± 5.0 %HR	± 1.9E+00 %HR