

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.75	0.750004	0.270	± 0.9	± 1.5E+00 mm					
0 a 0.8	0.800400	-0.130	± 0.9	± 1.5E+00 mm					
0 a 0.9	0.900300	-0.020	± 0.9	± 1.5E+00 mm					
0 a 1	1.000450	-0.170	± 1.0	± 1.5E+00 mm					
0 a 1.2	1.200450	-0.160	± 1.0	± 1.5E+00 mm					
0 a 1.4	1.400500	-0.200	± 1.1	± 1.5E+00 mm					
0 a 1.6	1.600230	0.080	± 1.2	± 1.5E+00 mm					
0 a 1.85	1.850200	0.120	± 1.3	± 1.5E+00 mm					
0 a 2	2.001200	-0.870	± 1.4	± 1.5E+00 mm					