

Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)	
50.3 mmHg	50 mmHg	-0.3 mmHg	± 5 mmHg	±	1.3E+00 mmHg
100.0 mmHg	100 mmHg	0.0 mmHg	± 5 mmHg	±	1.9E+00 mmHg
150.0 mmHg	150 mmHg	0.0 mmHg	± 5 mmHg	±	6.6E-01 mmHg
199.5 mmHg	200 mmHg	0.5 mmHg	± 5 mmHg	±	1.1E+00 mmHg
249.8 mmHg	250 mmHg	0.3 mmHg	± 5 mmHg	±	1.2E+00 mmHg
300.0 mmHg	300 mmHg	0.0 mmHg	± 5 mmHg	±	7.5E-01 mmHg