

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
50.1 mmHg	50 mmHg	-0.1 mmHg	± 5 mmHg	±	6.6E-01 mmHg
100.5 mmHg	100 mmHg	-0.5 mmHg	± 5 mmHg	±	1.0E+00 mmHg
149.5 mmHg	150 mmHg	0.5 mmHg	± 5 mmHg	±	8.7E-01 mmHg
200.3 mmHg	200 mmHg	-0.3 mmHg	± 5 mmHg	±	8.9E-01 mmHg
249.8 mmHg	250 mmHg	0.3 mmHg	± 5 mmHg	±	9.0E-01 mmHg
299.5 mmHg	300 mmHg	0.5 mmHg	± 5 mmHg	±	9.4E-01 mmHg