

Clase de Exactitud (Accuracy Class): Clase I (Class I)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	
m	m	mm	mm			m	m	mm	mm		
0 a 0.3	0.29981	0.190	± 0.2	±	4.3E-01 mm						
0 a 1	0.99991	0.090	± 0.4	±	4.3E-01 mm						
0 a 1.5	1.50008	-0.080	± 0.5	±	4.8E-01 mm						
0 a 2	1.99995	0.050	± 0.6	±	4.8E-01 mm						
0 a 2.5	2.49982	0.180	± 0.7	±	4.8E-01 mm						
0 a 3	2.99975	0.250	± 0.8	±	4.8E-01 mm						
0 a 3.5	3.49973	0.270	± 0.9	±	4.8E-01 mm						
0 a 4	3.99968	0.320	± 1	±	4.8E-01 mm						
0 a 4.5	4.49962	0.380	± 1.1	±	4.8E-01 mm						
0 a 5	4.99967	0.330	± 1.2	±	4.8E-01 mm						