

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
0.063 in	0.062 in	-0.001 in	± 0.002 in	±	7.0E-02 in
0.100 in	0.099 in	-0.001 in	± 0.002 in	±	7.6E-02 in
0.125 in	0.120 in	-0.005 in	± 0.002 in	±	7.6E-02 in
0.200 in	0.199 in	-0.001 in	± 0.002 in	±	9.4E-02 in
0.250 in	0.245 in	-0.005 in	± 0.002 in	±	7.0E-02 in
0.300 in	0.299 in	-0.001 in	± 0.002 in	±	9.9E-02 in
0.500 in	0.495 in	-0.005 in	± 0.002 in	±	7.6E-02 in
1.000 in	0.990 in	-0.010 in	± 0.002 in	±	1.0E-01 in