

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
10 °C	10.709 °C	10.7 °C	-0.009 °C	± 1.0 °C	± 4.0E-01 °C
20 °C	20.336 °C	20.4 °C	0.064 °C	± 5.0 °C	± 4.0E-01 °C
30 °C	30.363 °C	30.5 °C	0.137 °C	± 5.0 °C	± 4.0E-01 °C
40 °C	40.391 °C	40.6 °C	0.209 °C	± 5.0 °C	± 4.0E-01 °C