

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	0.060 °C	0.2 °C	0.140 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.274 °C	5.4 °C	0.126 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.238 °C	10.4 °C	0.162 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.220 %HR	31.8 %HR	2.580 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.118 %HR	52.0 %HR	2.882 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.897 %HR	73.9 %HR	3.003 %HR	± 3.0 %HR	± 1.9E+00 %HR