

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
12 °C	12.063 °C	12.2 °C	0.137 °C	± 0.5 °C	± 4.0E-01 °C
27 °C	27.235 °C	27.3 °C	0.065 °C	± 0.5 °C	± 4.0E-01 °C
42 °C	42.226 °C	42.3 °C	0.074 °C	± 0.5 °C	± 4.0E-01 °C
52 %HR	51.511 %HR	53.8 %HR	2.289 %HR	± 3.0 %HR	± 1.9E+00 %HR
72 %HR	71.807 %HR	74.2 %HR	2.393 %HR	± 3.0 %HR	± 1.9E+00 %HR
# 92 %HR	91.859 %HR	93.9 %HR	2.041 %HR	± 3.0 %HR	± 1.9E+00 %HR