

Certificado No. (Certificate #):		59709		Equipo (Instrument):		Medidor de Vibración			
Fecha de Calibración (Calibration Date):		2025-08-25		Marca (Brand):		HTI			
Patrón (Standard)		UBP (UUT)		Error (Error)		Exactitud (Accuracy)		Incertidumbre (Uncertainty)	
7 Hz	2.8 m/s^2-rms	2.7 m/s^2-rms	-0.1 m/s^2-rms			±	1.3E-01	m/s^2-rms	
10 Hz	5.5 m/s^2-rms	5.4 m/s^2-rms	-0.1 m/s^2-rms			±	1.6E-01	m/s^2-rms	
50 Hz	6.9 m/s^2-rms	7.0 m/s^2-rms	0.1 m/s^2-rms			±	1.8E-01	m/s^2-rms	
100 Hz	6.9 m/s^2-rms	6.9 m/s^2-rms	-0.1 m/s^2-rms			±	1.8E-01	m/s^2-rms	
300 Hz	6.9 m/s^2-rms	7.1 m/s^2-rms	0.1 m/s^2-rms			±	1.8E-01	m/s^2-rms	
500 Hz	6.9 m/s^2-rms	6.9 m/s^2-rms	0.0 m/s^2-rms			±	1.8E-01	m/s^2-rms	