

Patrón (Standard)		UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
7 Hz	0.551 m/s^2-rms	0.7 m/s^2-rms	0.15 m/s^2-rms		±	0.087 m/s^2-rms
7 Hz	1.523 m/s^2-rms	1.7 m/s^2-rms	0.17 m/s^2-rms		±	0.087 m/s^2-rms
7 Hz	2.771 m/s^2-rms	2.8 m/s^2-rms	0.03 m/s^2-rms		±	0.087 m/s^2-rms
10 Hz	3.043 m/s^2-rms	3.2 m/s^2-rms	0.15 m/s^2-rms		±	0.127 m/s^2-rms
10 Hz	5.543 m/s^2-rms	5.6 m/s^2-rms	0.06 m/s^2-rms		±	0.127 m/s^2-rms
30 Hz	6.936 m/s^2-rms	7 m/s^2-rms	0.06 m/s^2-rms		±	0.155 m/s^2-rms
50 Hz	6.936 m/s^2-rms	7.1 m/s^2-rms	0.16 m/s^2-rms		±	0.152 m/s^2-rms
80 Hz	6.936 m/s^2-rms	7 m/s^2-rms	0.06 m/s^2-rms		±	0.153 m/s^2-rms
100 Hz	6.936 m/s^2-rms	7.2 m/s^2-rms	0.26 m/s^2-rms		±	0.154 m/s^2-rms