

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.240 °C	-25.5 °C	-0.260 °C	± 1.0 °C	± 4.0E-01 °C
0 °C	0.080 °C	0.1 °C	0.020 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.495 °C	5.3 °C	-0.195 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.509 °C	10.4 °C	-0.109 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.349 °C	24.9 °C	-0.449 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.363 °C	30.0 °C	-0.363 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.491 °C	40.2 °C	-0.291 °C	± 0.5 °C	± 4.0E-01 °C

