

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| -25 °C            | -25.240 °C        | -24.8 °C  | 0.440 °C      | ± 1.0 °C             | ± 4.0E-01 °C                |
| 0 °C              | 0.080 °C          | 0.3 °C    | 0.220 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 5 °C              | 5.495 °C          | 5.2 °C    | -0.295 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 10 °C             | 10.509 °C         | 10.6 °C   | 0.091 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 25 °C             | 25.349 °C         | 25.6 °C   | 0.251 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 30 °C             | 30.363 °C         | 30.3 °C   | -0.063 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 40 °C             | 40.491 °C         | 40.4 °C   | -0.091 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |

