

Clase de Exactitud (Accuracy Class): Clase I (Class I)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.05	0.04985	0.150		± 4.3E-01 mm					
0 a 0.1	0.09992	0.080		± 4.3E-01 mm					
0 a 0.15	0.14988	0.120		± 4.3E-01 mm					
0 a 0.5	0.49974	0.260		± 4.3E-01 mm					
0 a 1	0.99973	0.270		± 4.3E-01 mm					
0 a 2	2.00044	-0.440		± 4.8E-01 mm					
0 a 3	2.99968	0.320		± 4.8E-01 mm					
0 a 4	4.00032	-0.320		± 4.8E-01 mm					
0 a 5	4.99973	0.270		± 4.8E-01 mm					
0 a 6	5.99972	0.280		± 3.6E-01 mm					
0 a 7	7.00015	0.790		± 3.6E-01 mm					
0 a 8	8.00095	0.180		± 3.6E-01 mm					
0 a 9	9.00190	-0.550		± 3.6E-01 mm					
0 a 10	9.99930	0.698		± 3.6E-01 mm					
0 a 11	11.00012	-0.123		± 3.6E-01 mm					
0 a 12	12.00013	-0.133		± 3.6E-01 mm					
0 a 13	13.00064	-0.642		± 3.6E-01 mm					
0 a 14	14.00015	-0.145		± 3.6E-01 mm					
0 a 15	14.99974	0.262		± 3.6E-01 mm					
0 a 16	16.00018	-0.179		± 3.6E-01 mm					
0 a 17	16.99933	0.674		± 3.6E-01 mm					
0 a 18	18.00038	-0.376		± 3.6E-01 mm					
0 a 19	18.99929	0.714		± 3.6E-01 mm					
0 a 20	20.00014	-0.137		± 3.6E-01 mm					