

Certificado No. (Certificate #):		59754	Equipo (Instrument):		Reloj Palpador
Fecha de Calibración (Calibration Date):		2025-08-08	Marca (Brand):		Mitutoyo
Patrón (Standard)		UBP (UUT)	Error (Error)	Exactitud (Accuracy)	
0.050 in		0.050 in	0.000 in	± 6.8E-05 in	
0.100 in		0.100 in	0.000 in	± 9.1E-05 in	
0.110 in		0.110 in	0.000 in	± 9.6E-05 in	
0.130 in		0.130 in	0.000 in	± 1.1E-04 in	
0.150 in		0.150 in	0.000 in	± 1.2E-04 in	
0.180 in		0.180 in	0.000 in	± 1.4E-04 in	
0.200 in		0.200 in	0.000 in	± 1.5E-04 in	
0.300 in		0.300 in	0.000 in	± 2.2E-04 in	
0.500 in		0.500 in	0.000 in	± 3.5E-04 in	