

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.463 °C	30.1 °C	-0.363 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.059 %HR	19.0 %HR	-11.059 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.529 %HR	39.0 %HR	-11.529 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	61.0 %HR	-9.448 %HR	± 5.0 %HR	± 1.9E+00 %HR