

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.463 °C	30.6 °C	0.137 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.059 %HR	31.0 %HR	0.941 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.529 %HR	47.0 %HR	-3.529 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	67.0 %HR	-3.448 %HR	± 5.0 %HR	± 1.9E+00 %HR