

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
50.60 μm	49.80 μm	-0.80 μm	± 2 μm	±	5.8E-01 μm
120.30 μm	120.6 μm	0.30 μm	± 2 μm	±	5.8E-01 μm
246.20 μm	244.6 μm	-1.60 μm	± 2 μm	±	5.8E-01 μm
479.00 μm	479.4 μm	0.40 μm	± 2 μm	±	5.8E-01 μm