

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
52.7 μm	52.7 μm	0.05 μm	± 5 μm	±	5.8E-01 μm
155.2 μm	154.9 μm	-0.33 μm	± 5 μm	±	5.8E-01 μm
257.12 μm	257.40 μm	0.28 μm	± 5 μm	±	5.8E-01 μm
511.72 μm	512.00 μm	0.28 μm	± 5 μm	±	5.8E-01 μm