

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.393 °C	29.9 °C	-0.493 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.376 %HR	30 %HR	-0.376 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.161 %HR	48 %HR	-2.161 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	73.904 %HR	72 %HR	-1.904 %HR	± 5.0 %HR	± 1.9E+00 %HR