

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	
m	m	mm	mm			m	m	mm	mm		
0 a 0.5	0.49928	0.720	± 0.8	±	4.3E-01 mm						
0 a 1	0.99918	0.820	± 1	±	4.3E-01 mm						
0 a 1.5	1.49931	0.690	± 1.2	±	4.8E-01 mm						
0 a 2	1.99938	0.620	± 1.4	±	4.8E-01 mm						
0 a 2.5	2.49908	0.920	± 1.6	±	4.8E-01 mm						
0 a 3	2.99903	0.970	± 1.8	±	4.8E-01 mm						
0 a 3.5	3.49910	0.900	± 2	±	4.8E-01 mm						
0 a 4	3.99905	0.950	± 2.2	±	4.8E-01 mm						
0 a 4.5	4.49967	0.330	± 2.4	±	4.8E-01 mm						
0 a 5	4.99902	0.980	± 2.6	±	4.8E-01 mm						