

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| 30 °C             | 30.463 °C         | 30.9 °C   | 0.437 °C      | ± 1.0 °C             | ± 4.0E-01 °C                |
| 30 %HR            | 30.028 %HR        | 28.0 %HR  | -2.028 %HR    | ± 5.0 %HR            | ± 1.9E+00 %HR               |
| 50 %HR            | 50.529 %HR        | 47.0 %HR  | -3.529 %HR    | ± 5.0 %HR            | ± 1.9E+00 %HR               |
| 70 %HR            | 70.039 %HR        | 72.0 %HR  | 1.961 %HR     | ± 5.0 %HR            | ± 1.9E+00 %HR               |