

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.463 °C	31.3 °C	0.837 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.028 %HR	26.0 %HR	-4.028 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.529 %HR	47.0 %HR	-3.529 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.039 %HR	70.0 %HR	-0.039 %HR	± 5.0 %HR	± 1.9E+00 %HR