

| Patrón (Standard) | UBP (UUT) | Error (Error) | E.M.P (Tolerance) | Incertidumbre (Uncertainty) |
|--------------------------|---------------------|----------------------|--------------------------|------------------------------------|
| 52.65 μm | 52.82 μm | 0.17 μm | $\pm 2.5 \mu\text{m}$ | $\pm 5.8\text{E-}02 \mu\text{m}$ |
| 155.23 μm | 155.6 μm | 0.37 μm | $\pm 2.5 \mu\text{m}$ | $\pm 6.2\text{E-}02 \mu\text{m}$ |
| 257.12 μm | 257.2 μm | 0.08 μm | $\pm 2.5 \mu\text{m}$ | $\pm 7.0\text{E-}02 \mu\text{m}$ |
| 511.71 μm | 511.8 μm | 0.09 μm | $\pm 2.5 \mu\text{m}$ | $\pm 9.7\text{E-}02 \mu\text{m}$ |