

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.141 °C	15.1 °C	-0.041 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.291 °C	39.9 °C	-0.391 °C	± 0.5 °C	± 4.0E-01 °C
25 %HR	25.151 %HR	25.3 %HR	0.149 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.039 %HR	70.7 %HR	0.661 %HR	± 5.0 %HR	± 1.9E+00 %HR