

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(M.P.E)</i>	Incertidumbre <i>(Uncertainty)</i>	
50.1039 ml	50.0 ml	-0.104 ml	± 2.5 ml	±	3.5E-01 µl
100.2429 ml	100.0 ml	-0.243 ml	± 2.5 ml	±	3.5E-01 µl
500.0205 ml	500.0 ml	-0.020 ml	± 2.5 ml	±	3.7E-01 µl