

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
60 RPM	59.999 RPM	-0.001 RPM	± 0.026 RPM	±	4.0E-03 RPM
600 RPM	599.99 RPM	-0.01 RPM	± 0.26 RPM	±	7.0E-03 RPM
6000 RPM	5999.9 RPM	-0.1 RPM	± 2.6 RPM	±	5.8E-02 RPM
60000 RPM	60000 RPM	0 RPM	± 26 RPM	±	8.0E-01 RPM
99960 RPM	99951 RPM	-9 RPM	± 42 RPM	±	8.4E-01 RPM