

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	
m	m	mm	mm	(Uncertainty)		m	m	mm	mm	(Uncertainty)	
0 a 0.01	0.00990	0.100	± 0.6	±	4.3E-01 mm						
0 a 0.02	0.01978	0.220	± 0.6	±	4.3E-01 mm						
0 a 0.03	0.02991	0.090	± 0.6	±	4.3E-01 mm						
0 a 0.04	0.03981	0.190	± 0.6	±	4.3E-01 mm						
0 a 0.05	0.04988	0.120	± 0.6	±	4.3E-01 mm						
0 a 0.06	0.05995	0.050	± 0.6	±	4.3E-01 mm						
0 a 0.07	0.06995	0.050	± 0.6	±	4.3E-01 mm						
0 a 0.08	0.07993	0.070	± 0.6	±	4.3E-01 mm						
0 a 0.09	0.09011	-0.110	± 0.6	±	4.3E-01 mm						
0 a 0.1	0.09998	0.020	± 0.6	±	4.3E-01 mm						
0 a 0.11	0.11021	-0.210	± 0.6	±	4.3E-01 mm						